



## PRODUCT ANNOUNCEMENT LETTER

### Interlaboratory Quality Assurance Program (IQAP)

Dear Beckman Coulter Customer,

We are pleased to announce updates to the Beckman Coulter eIQAP website.

This letter provides a summary of changes/updates in the eIQAP website. Please visit the Beckman Coulter website at <https://qapweb.beckmancoulter.com/qapweb/> to access the latest Instructions for Use.

#### **IQAP Website Changes for v2.3**

##### **1. Features**

General Changes	<ul style="list-style-type: none"><li>• All links and information referencing the IQAP chemistry website were removed from eIQAP website.</li><li>• Corrected "Privacy Policy" link on the footer of the website.</li></ul>
Single Sign On	<p>The following changes will be noticed:</p> <ul style="list-style-type: none"><li>• On the eIQAP home page, the current Sign In screen which includes the User ID and Password fields has been replaced with a Sign In&gt;Log In Single Sign On (SSO).</li><li>• On the current Sign In screen, the <b>Forgot username or password?</b> link has been removed and is now part of the SSO login page.</li><li>• The current My Profile screen includes information pertaining to the user and institution. The new My Profile screen only includes the information pertaining to the user.</li><li>• The current Account Registration screen includes information pertaining to User, Institution, and a Hint Question. The new Account Registration screen only includes the information pertaining to the User.</li><li>• The current Sign Out screen states <b>You have signed out</b> and has an option to <b>Return to the Dashboard</b>. The new Sign Out screen states the user with <b>You have logged out</b> and will momentarily route the User back to the Home screen.</li></ul>



## 2. Bug Fixes

- a. Customers will be able to manually type in a date when entering control data.
- b. Minimum pool sizes will be met when a pool size is greater than minimum on the 6C Plus Control report.
- c. When uploading Linearity data on Hematology instruments, the “pending with errors” message is fixed
- d. Fixed the report bug, when report status on the final reports will read “reported” instead of “pending”.

Thank you for choosing Beckman Coulter products for your Hematology laboratory and Life Science choices, and your continued support of Beckman Coulter products.

Sincerely,

A handwritten signature in black ink, appearing to read 'Pingping Lou'.

Pingping Lou

Sr. Marketing Management, Global Product Management